

Potassium Chloride, Saturated, approximately 4 Molar

Lot Number: 4402B86

Product Number: 5900

Manufacture Date: JAN 31, 2024

Expiration Date: JAN 2026

All lots of this product are certified suitable for the following applications: 1. For refilling Calomel type Combination pH Reference Electrodes; 2. For refilling Calomel type pH Reference Half Cell Electrodes; 3. For refilling some brands of Silver/Silver Chloride Combination pH or pH Reference Half Cell Electrodes with a Silver Chloride Internal Reference that is sealed by an ion exchange barrier; 4. For use as an Ionic Strength Adjustor for some brands of Calcium Ion Selective Electrodes; 5. For refilling some brands of Single Junction Calcium, Single Junction Fluoride, and Combination Sodium Ion Selective Electrodes. This product can also be used for other applications where the concentration and matrix are appropriate.

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Potassium Chloride	7447-40-7	ACS
Test	Specification	Result
Appearance	Colorless liquid	Passed
Specification	Reference	
Potassium Chloride Electrolyte		ASTM (D 664)
Potassium Chloride Electrolyte		ASTM (D 974)
Potassium Chloride Electrolyte	د	ASTM (D 3242)
Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.		
Part Number Size	/ Package Type	Shelf Life (Unopened Container)

Recommended Storage: 15°C - 30°C (59°F - 86°F)

Yand Dran

Paul Brandon (01/31/2024) Production Manager This document is designed to comply with ISO Guide 31 "Reference Materials --Contents of Certificates and Labels."

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.

4 L natural poly

500 mL natural poly

5900 - 1

5900-16

24 months

24 months