1490 Lammers Pike Batesville, IN 47006 http://www.riccachemical.com 1-888-GO-RICCA

customerservice@riccachemical.com

## Certificate of Analysis

Potassium Permanganate, 1.00 Normal

Lot Number: 4402C40 Product Number: 6420

Manufacture Date: FEB 02, 2024

Expiration Date: JUL 2025

Name	CAS#	Grade	
Water	7732-18-5	ACS/ASTM/USP/EP	
Potassium Permanganate	7722-64-7	ACS	

Test	Specification	Result	NIST SRM#
Appearance	Purple liquid	Passed	
Assay (vs. Sodium Oxalate)	0.999-1.001 N at 20°C	1.000 N at 20°C	8040

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
6420-1	4 L amber glass	18 months
6420-16	500 mL amber glass	18 months
6420-32	1 L amber glass	18 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)

Paul Brandon (02/02/2024)

**Production Manager** 

This document is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.

Version: 1.4 Lot Number: 4402C40 Product Number: 6420 Page 1 of 1