

Certificate of Analysis

Phosphate Buffer Solution, pH 6.8, for Phenols

Lot Number: 4402P77

Product Number: 5807

Manufacture Date: FEB 29, 2024

Expiration Date: FEB 2026

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Potassium Phosphate, Dibasic	7758-11-4	ACS
Potassium Dihydrogen Phosphate	7778-77-0	ACS

Test	Specification	Result	NIST SRM#
Appearance	Colorless liquid	Passed	
pH at 25°C (Method: SQCP027, SQCP033)	6.75-6.85	6.77	186-I-g, 186-II-g, 191d

Specification	Reference
Phosphate Buffer Solution	APHA (5530 C)
Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.	

Part Number	Size / Package Type	Shelf Life (Unopened Container)
5807-16	500 mL natural poly	24 months
5807-32	1 L natural poly	24 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)



Paul Brandon (02/29/2024)

Production Manager

This document is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.