

Certificate of Analysis

Perchloric Acid, 0.5 N in Ethylene Glycol

Lot Number: 4404P95

Product Number: R5516500

Manufacture Date: APR 30, 2024

Expiration Date: APR 2026

Name	CAS#	Grade
Water	7732-18-5	ACS
Perchloric Acid	7601-90-3	ACS
Ethylene Glycol	107-21-1	Reagent

Test	Specification	Result	NIST SRM#
Appearance	clear, colorless	Passed	
Assay (vs. Sodium Hydroxide/Thymol Blue)	0.495-0.505	0.500 N at 20°C	84

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
R5516500-1C	1 L amber glass	24 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)



Paul Brandon (04/30/2024)

Production Manager

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.