

Certificate of Analysis

Ferrous Ammonium Sulfate, 0.100 Normal (N/10), 0.100 Molar (M/10)

Lot Number: 4405C01

Product Number: 3145

Manufacture Date: MAY 02, 2024

Expiration Date: APR 2025

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Sulfuric Acid	7664-93-9	ACS
Ferrous Ammonium Sulfate Hexahydrate	7783-85-9	ACS

Test	Specification	Result	NIST SRM#
Appearance	Light green liquid	Passed	
Assay (vs. Potassium Dichromate/Ferrouin)	0.0999-0.1001 N at 20°C	0.1000 N at 20°C	136

Specification	Reference
Ferrous Ammonium Sulfate Solution (0.10 N)	ASTM (D 1252 B)
Standard Ferrous Ammonium Sulfate Titrant (FAS)	APHA (5220 C)
Ferrous ammonium sulfate, 0.1 N	TAPPI (T 610 sp-97)
Ferrous Ammonium Sulfate, 0.1 Normal Volumetric Solution	ACS (N/A)

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
3145-1	4 L amber glass	12 months
3145-16	500 mL amber glass	12 months
3145-32	1 L amber glass	12 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)



Paul Brandon (05/02/2024)

Production Manager

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.