

# Certificate of Analysis

**Methylene Blue, 0.1% (w/v)****Lot Number:** 4406E42**Product Number:** 4841**Manufacture Date:** JUN 10, 2024**Expiration Date:** MAY 2026

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Methylene Blue Trihydrate	7220-79-3	Certified

Test	Specification	Result
Appearance	Blue liquid	Passed
UV/Visible Spectrum	To Pass Test	Passed

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
4841-1	4 L amber glass	24 months

**Recommended Storage:** 15°C - 30°C (59°F - 86°F)Paul Brandon (06/10/2024)  
Production Manager

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.