Arlington, TX 76012
Pocomoke City, MD 21851
Batesville, IN 47006
http://www.riccachemical.com
1-888-GO-RICCA
customerservice@riccachemical.com

Certificate of Analysis

HydroSpec® Coulometric AG

General Purpose Anolyte Solution for Cells With or Without Diaphragm

Lot Number: Q307R08 Product Number: RK100000

Manufacture Date: FEB 27, 2024

Recommended Retest Date: JUL 2028

Name	CAS#	Grade	
Methyl Alcohol	67-56-1		
Sulfur Dioxide	7446-09-5		
Imidazole	288-32-4		
Diethanolamine	111-42-2		
Imidazolium Iodide	68007-08-9		

Test	Specification	Result
Suitability for Use	Karl Fischer titration	Passed
Water (H_2O)	0.002-0.030 %	0.004 %

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
RK100000-500C	500 mL amber glass	60 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)

Paul Brandon (02/27/2024)

Production Manager

This document is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.

Version: 1.2 Lot Number: Q307R08 Product Number: RK100000 Page 1 of 1